



**Vermont Department of Health**  
**Division of Health Surveillance**  
Public Health Laboratory  
195 Colchester Avenue - P.O. Box 1125  
Burlington, VT 05402-1125

[phone] 802-863-7335  
[fax] 802-863-7632  
[or] 800-660-9997

*Agency of Human Services*

**COUNTY: Addison**

Page 1 of 1

July 14, 2011

Dear colleague:

Enclosed is a packet of information that includes all technical information regarding each of the active DataMaster DMT instruments in your county. This documentation would include maintenance and repair records, certification records, and installation records. This would include documents created or received during June 2011.

If you have questions or suggestions regarding this process, please don't hesitate to contact me at (802) 863-7570.

Thank you.

Sincerely,

A handwritten signature in cursive script, appearing to read "Mary Celotti".

Mary Celotti  
Laboratory Director

Enclosures:

**Datamaster DMT S.N. 100137 (Middlebury Police Dept)**

Routine Performance Check Report, 06/09/11

**Datamaster DMT S.N. 100135 (New Haven VSP)**

No updates on technical information during June 2011

**Datamaster DMT S.N. 100141 (Vergennes Police Dept)**

Routine Performance Check Report with 1 page attachment, 06/21/11

**Datamaster DMT S.N. 100134 (Bristol Police Dept)**

No updates on technical information during June 2011




RECEIVED

ROUTINE PERFORMANCE CHECK REPORT

JUN 14 2011

DataMaster DMT: 100137  
Location: MIDDLEBURY PD  
Calibration Date: 04/27/2010  
Certification Date: 04/27/2010  
Installation Date: 04/28/2010  
RPC Date: 06/09/2011  
Supervisor Name: CHRISTOPHER L MASON



Diagnostic Results *SK*

VERSIONS  
DMT 1.05  
PIC 2.05  
Modem 1.06  
Questions 1.00  
Reports 1.02

TEMPERATURES *SK*  
Sample Chamber = 48.86°C  
Breath Tube = 47.79°C  
Digital Sim = 34.1°C

SETTINGS *SK*  
Lamp Voltage = 1.59 V  
Cooler Voltage = 1.66 V  
Bias Voltage = 79.9 V  
Chopper Freq = 550.0 Hz

PUMP INFO *SK*  
Flow Rate = 5.333 L/M

DETECTOR INFO *SK*  
PUMP ON OFF  
MAX(V) 0.0010 0.0055  
MIN(V) -0.0010 0.0028

FILTER INFO *SK*  
Filter 1 0.004 Zero = true  
Filter 2 0.282 Zero = true  
Filter 3 0.473 Zero = true

CALIBRATION CHECK *SK*  
Xq = 0.0820 1.15%

Routine Performance Check Passed *SK*

Accuracy and Precision Check *SK*


Concentration = 0.097 g/210L  
Lot # = 11-20-100  
Average = 0.101 g/210L  
Std Dev = 0.001

RFI Interference Test *SK*

RFI detected

Sample Acceptance Test *SK*

Passed



Performed by *CH*  
Reviewed by *Steven Harnois*

Date 06/09/2011

Date 6/17/11

DMT Serial Number #100137

Page 1 of 1  
DMT 100137-000073

Reviewed *KKK* 6-16-11

*6/16/11*

*6/20/11*

06/09/2011 23:10